DECLARATION OF CONFORMITY

We are solely responsible for declaring that the Medical Devices mentioned in this statement are of Low-Risk Class (Class I) and comply with the requirements of the European Regulation 2017/745 and where appropriate, the standards and legislation referred to.

MANUFACTURER:	**MOBIAK S.A**
SRN:	**GR-MF-000016256**
SEAT ADDRESS:	**KATHIANA AKROTIRIOU-CHANIA-CRETE-GREECE**
E-MAIL:	**mobiakcarequality@mobiak.com**
TELEPHONE:	**+30 2821063222**
WEB SITE:	**www.mobiakcare.com**

COMPETENT AUTHORITY:	**National Organization for Medicines**
CONTACT DETAILS:	**Address: Mesogeion 284, PS 15562, Cholargos** **Phone : 2132040000** **Website: https://www.eof.gr/**

LIST OF PRODUCTS COVERED BY THIS DECLARATION				
PRODUCT	**CODE**	**UDI-DI**	**INTENDED USE**	**RULE**
PATIENT TRANSFER STRETCHER	0806458	521300690STRETCHERS184Z	IT IS SUITABLE FOR TRANSPORTATION OF PATIENTS IN EMERGENCY CONDITIONS OR IN CONDITIONS THAT REQUIRE QUICK AND SAFE REMOVAL OF THE PATIENT.	1
SOFT STRETCHER	0806459	521300690STRETCHERS184Z	IT IS SUITABLE FOR TRANSPORTATION OF PATIENTS IN EMERGENCY CONDITIONS OR IN CONDITIONS THAT REQUIRE QUICK AND SAFE REMOVAL OF THE PATIENT.	1
FOLDING STRETCHER, SINGLE FOLD	0806460	521300690STRETCHERS184Z	IT IS SUITABLE FOR TRANSPORTATION OF PATIENTS IN EMERGENCY CONDITIONS OR IN CONDITIONS THAT REQUIRE QUICK AND SAFE REMOVAL OF THE PATIENT.	1
FOLDING STRETCHER, SINGLE FOLD	0804591	521300690STRETCHERS184Z	IT IS SUITABLE FOR TRANSPORTATION OF PATIENTS IN EMERGENCY CONDITIONS OR IN CONDITIONS THAT REQUIRE QUICK AND SAFE REMOVAL OF THE PATIENT.	1
STAIR CHAIR STRETCHER	0804593	521300690STRETCHERS184Z	IT IS SUITABLE FOR TRANSPORTATION OF PATIENTS IN EMERGENCY CONDITIONS OR IN CONDITIONS THAT REQUIRE QUICK AND SAFE REMOVAL OF THE PATIENT.	1
SOFT STRETCHER	0804592	521300690STRETCHERS184Z	IT IS SUITABLE FOR TRANSPORTATION OF PATIENTS IN EMERGENCY CONDITIONS OR IN CONDITIONS THAT REQUIRE QUICK AND SAFE REMOVAL OF THE PATIENT.	1

BASKET STRETCHER	0805906	521300690STRETCHERS184Z	IT IS SUITABLE FOR TRANSPORTATION OF PATIENTS IN EMERGENCY CONDITIONS OR IN CONDITIONS THAT REQUIRE QUICK AND SAFE REMOVAL OF THE PATIENT.	1
SPINE BOARD	0805907	521300690STRETCHERS184Z	IT IS SUITABLE FOR TRANSPORTATION OF PATIENTS IN EMERGENCY CONDITIONS OR IN CONDITIONS THAT REQUIRE QUICK AND SAFE REMOVAL OF THE PATIENT.	1
SCOOP STRETCHER	0805910	521300690STRETCHERS184Z	IT IS SUITABLE FOR TRANSPORTATION OF PATIENTS IN EMERGENCY CONDITIONS OR IN CONDITIONS THAT REQUIRE QUICK AND SAFE REMOVAL OF THE PATIENT.	1
FOLDING STRETCHER, SINGLE FOLD	0804591	521300690STRETCHERS184Z	IT IS SUITABLE FOR TRANSPORTATION OF PATIENTS IN EMERGENCY CONDITIONS OR IN CONDITIONS THAT REQUIRE QUICK AND SAFE REMOVAL OF THE PATIENT.	1
SOFT STRETCHER	0804592	521300690STRETCHERS184Z	IT IS SUITABLE FOR TRANSPORTATION OF PATIENTS IN EMERGENCY CONDITIONS OR IN CONDITIONS THAT REQUIRE QUICK AND SAFE REMOVAL OF THE PATIENT.	1
STAIR CHAIR STRETCHER	0804593	521300690STRETCHERS184Z	IT IS SUITABLE FOR TRANSPORTATION OF PATIENTS IN EMERGENCY CONDITIONS OR IN CONDITIONS THAT REQUIRE QUICK AND SAFE REMOVAL OF THE PATIENT.	1

FOLDING STRETCHER WITH WHEELS	0805918	521300690STRETCHERS184Z	IT IS SUITABLE FOR TRANSPORTATION OF PATIENTS IN EMERGENCY CONDITIONS OR IN CONDITIONS THAT REQUIRE QUICK AND SAFE REMOVAL OF THE PATIENT.	1
TRANSPORT STRETCHER CHAIR (2 IN 1)	0805919	521300690STRETCHERS184Z	IT IS SUITABLE FOR TRANSPORTATION OF PATIENTS IN EMERGENCY CONDITIONS OR IN CONDITIONS THAT REQUIRE QUICK AND SAFE REMOVAL OF THE PATIENT.	1
STAIR CHAIR CLIMBER	0805976	521300690stretchers184Z	IT IS SUITABLE FOR TRANSPORTING PATIENTS IN EMERGENCY SITUATIONS OR IN SITUATIONS THAT REQUIRE QUICK AND SAFE REMOVAL OF THE PATIENT.	1
HOSPITAL STRETCHER WITH WHEELS	0805945	521300690stretchers184Z	IT IS SUITABLE FOR TRANSPORTING PATIENTS IN EMERGENCY SITUATIONS OR IN SITUATIONS THAT REQUIRE QUICK AND SAFE REMOVAL OF THE PATIENT.	1
ELECTRIC STAIR CLIMBING SYSTEM FOR WHEELCHAIR	0807109	521300690stretchers184Z	USED TO TRANSPORT PEOPLE WITH LIMITED ABILITY TO CLIMB STAIRS SAFELY DUE TO PHYSICAL LIMITATIONS	1
ALUMINUM ELECTRIC STAIR CHAIR CLIMBER	0807110	521300690stretchers184Z	USED TO TRANSPORT PATIENTS UP STAIRS, IN EMERGENCY SITUATIONS OR IN SITUATIONS THAT REQUIRE QUICK AND SAFE REMOVAL OF THE PATIENT.	1
ALUMINUM ELECTRIC STAIR CHAIR CLIMBER	0805996	521300690stretchers184Z	IT IS SUITABLE FOR TRANSPORTING PATIENTS IN EMERGENCY SITUATIONS OR IN SITUATIONS THAT REQUIRE QUICK AND SAFE REMOVAL OF THE PATIENT.	1

STAINLESS STEEL PATIENT STRETCHER	0805948	521300690stretchers184Z	INDICATED FOR THE SAFE AND EFFICIENT TRANSPORT OF PATIENTS WITHIN HEALTHCARE FACILITIES AND DURING EMERGENCY AND URGENT CARE SITUATIONS	1

CONFORMITY ASSESSMENT PROCEDURE
According to Annexes II & III of Regulation (EU) 2017/745

APPLIED STANDARDS & LEGAL REQUIREMENTS
ISO 13485:2016, ISO 9001:2015, EN ISO 15223-1: 2016, EN ISO 14971:2019, EN ISO 10993-1:2018, EN ISO 10993-1:2009, ISO 10993-10:2021 , (EU) 2017/745

CE

FOR APPROVAL	
NAME:	**SVOURAKI MARIA**
POSITION:	**CEO**
PLACE:	**CHANIA**
DATE:	**10/10/2025**
SIGN:	

